

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 07 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130492 ✓
Name <u>Shayne Calt</u>		ID# <u>4559</u>	Date <u>3-15-09</u> ✓
A	Agency <u>AST</u>	Phone #	<u>733-2256</u>
Instrument Location <u>Willow AST</u>			
Alco S/N <u>X301581</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>1200</u>
B	Alco Test Value Average		
	<u>.082</u> ✓ 1 st Alco	<u>.081</u> ✓ 2 nd Alco	
Signature <u>[Signature]</u>		<u>Calt</u> <u>7/14/09</u>	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of July, 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130492

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

MARCH 15, 2009

OPERATOR'S NAME:

CALT SHAYNE M

OPERATOR'S NUMBER: 4559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

V

D.L. #: V

DEPT/AGENCY: TALB

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .002

ALCO S/N: 430492 X 301581

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.29 in

ALCO TARGET .000 17:37

BLANK TEST .000 17:38

INTERNAL STANDARD VERIFIED 17:38

ALCO TV 29.29 in .002 17:38

BLANK TEST .000 17:39

SUBJECT SAMPLE .000 17:40

BLANK TEST .000 17:40

ALCO TV 29.31 in .001 17:41

BLANK TEST .000 17:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

MARCH 15, 2009
TIME 17:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 29.29 in

FLOW DETECTOR: OKAY

PUMP

WTH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
pqrstuvwxyz{!}~